


Certificate of Analysis

Polymer type: Poly(styrene-methyl methacrylate)

Part No: PSS-psmm51k

Lot No: psmm061006

Elugramm



GPC/SEC – Conditions

Sample concentration	1,00 g/l	Inject volume	20 µl
Solvent	Tetrahydrofuran	Flow rate	1.00 ml/min
Precolumn [8 x 50 mm]	PSS SDV 5µm		
Columns [analytical, each 8 x 300 mm]	PSS SDV 5µm 10e3Å / 10e5Å / 10e6Å		
Temperature	23 °C		
Data Acquisition Software	PSS WinGPC	Operator	R. Leinweber

GPC/SEC – Results*

Detector	Mw (total) [Da]	Mn (total) [Da]	Mp (total) [Da]	PDI (total) [Mw/Mn]
Agilent RID	50 700	34 200	50 700	1.49

* molar mass calculation based on Poly(styrene) calibration

Copolymer - Composition	[w %]
Poly(styrene)	83
Poly(methyl methacrylate)	17


Note:

Mw = Weight Average Molecular Weight
Mn = Number Average Molecular Weight
Mp = Molar Mass at the Peak Maximum
PDI = Polydispersity Index

6WRUD# Store the tightly recapped polymer standard in a dry, dark, cool area; e.g. a refrigerator (4 °C).

R EU PP GG 6ee al o product label.
a o approval

Manufacture and control according to PSS method of analysis


Dr. T. Hofe
production director