


Certificate of Analysis

Polymer type: Poly(styrene sulfonate-d8) sodium salt
 Part No: PSS-pssd1.5m
 Lot No: pssde240605

Molar Mass Distribution



GPC/SEC - Conditions

Sample concentration	0,10 g/l	Inject volume	100 µl
Flow rate	0,50 ml/min	Temperature	23 °C
Solvent	Water, Disodium hydrogen phosphate 11,88g/L		
Precolumn [8 x 50 mm]	PSS MCX 10µm		
Columns [analytical, each 8 x 300 mm]	PSS MCX 10µm 10e7Å		
Data Acquisition Software	PSS WinGPC	Operator	J.Preis

GPC/SEC - Results

Detector	Mw [Da]	Mn [Da]	Mp [Da]	PDI [Mw/Mn]
Agilent VWD 254nm	1500000	-	1580000	<1.20

Parent Poly(styrene-d8) Molecular Weight: Mw [Da] = 828000 Mn [Da] = 716000 Mp [Da] = 874000 PDI = 1.16

Note:

Mw = Weight average molecular weight
 Mn = Number average molecular weight
 Mp = Molar mass at the peak maximum
 PDI = Polydispersity Index

The molecular weights are calculated with the factor 1.81 (pssd sodium salt / psd x 0.95). Degree of sulfonation > 90%.
 (For calculation: Assumption: Degree of sulfonation is 95%).

Storage: Store the tightly recapped polymer standard in a dry, dark, cool area; e.g. a refrigerator (4 °C).

Date of expiry: See product label.

Manufacture and control according to PSS method of analysis



Dr. T. Hofe
production director

